



P.O. BOX 265
OGDENSBURG, NY
U.S.A. 13669-0265

TEL: 888-670-8729 (USA & Canada) or +1-613-226-5772 (Intl)
FAX: 800-561-1970 (USA & Canada) or +1-613-226-2802 (Intl)


BOX 5120, LCD MERIVALE
OTTAWA, ONTARIO
CANADA K2C 3H4

info@avtechpulse.com - http://www.avtechpulse.com/

PERFORMANCE CHECKSHEET

Model: AVRZ-5W-B-PN
Type: High-Speed, High-Voltage Pulse Generator
S.N.: 11416 (modification)
Date: April 29, 2008


Rise/Fall Time Test



29 Apr 2008 12:29:39
+500V and -500V, 50Ω load, 500 ns PW, +100 ns delay, 100 Hz. 200 V/div, 100 ns/div.

- a) Output Signal Amplitude: 0 to ±500V, to 50Ω
- b) Pulse Width: 15 ns - 10 us, 0.1% duty cycle
- c) Rise Time (20%-80%): < 10 ns (positive output)
- d) Fall Time (80%-20%): < 5 ns (positive output)
- e) PRF: 1 Hz - 5 kHz
- f) Jitter, Stability: OK
- g) Prime Power: 100-240V AC, 50-60 Hz


Maximum Duty Cycle Test



29 Apr 2008 12:31:30
+500V and -500V, 50Ω load, 10 us PW, -1 us delay, 100 Hz. 200 V/div, 400 ns/div.

Reference levels: 20%, 80%.

Short Circuit Test



29 Apr 2008 12:40:01
500 ns, -500V into a 50Ω load shorted with a 12" length of patch cord. 100 V/div, 100 ns/div.