



P.O. BOX 265  
OGDENSBURG, NY  
U.S.A. 13669-0265

TEL: 888-670-8729 (USA & Canada) or +1-613-226-5772 (Intl)  
FAX: 800-561-1970 (USA & Canada) or +1-613-226-2802 (Intl)


BOX 5120, LCD MERIVALE  
OTTAWA, ONTARIO  
CANADA K2C 3H4

info@avtechpulse.com - http://www.avtechpulse.com/

PERFORMANCE CHECKSHEET

Model: AVRZ-5W-B-PN  
Type: High-Speed, High-Voltage Pulse Generator  
S.N.: 12022  
Date: July 11, 2008

Rise/Fall Time Test



+500V, 50Ω load, PW = 500 ns.  
100 V/div, 100 ns/div.


11 Jul 2008  
08:47:48

- a) Output Signal Amplitude: 0 to ±500V, to 50Ω
- b) Pulse Width: 15 ns - 10 us, 0.1% duty cycle
- c) Rise Time (20%-80%): < 10 ns (positive output)
- d) Fall Time (80%-20%): < 10 ns (positive output)
- e) PRF: 1 Hz - 5 kHz

f) Jitter, Stability: OK

g) Prime Power: 100-240V AC, 50-60 Hz


Maximum Duty Cycle Test



-500V, 50Ω load, PW = 10 us, PRF = 100 Hz.  
100 V/div, 2 us/div.

11 Jul 2008  
08:48:42

Short Circuit Test



500 ns, -500V into a 50Ω load shorted with a 12" length of patch cord. 100 V/div, 100 ns/div.

11 Jul 2008  
08:50:35

Reference levels: 20%, 80%.