INSTRUCTIONS

MODEL AV-149-BW4-2.8K-PIN TRANSIMPEDANCE AMPLIFIER

S.N.:

<u>WARRANTY</u>

Avtech Electrosystems Ltd. warrants products of its manufacture to be free from defects in material and workmanship under conditions of normal use. If, within one year after delivery to the original owner, and after prepaid return by the original owner, this Avtech product is found to be defective, Avtech shall at its option repair or replace said defective item. This warranty does not apply to units which have been dissembled, modified or subjected to conditions exceeding the applicable specifications or ratings. This warranty is the extent of the obligation assumed by Avtech with respect to this product and no other warranty or guarantee is either expressed or implied.

TECHNICAL SUPPORT

Phone: 613-226-5772 or 1-800-265-6681 Fax: 613-226-2802 or 1-800-561-1970


E-mail: info@avtechpulse.com World Wide Web: http://www.avtechpulse.com

TABLE OF CONTENTS

WARRANTY2
TABLE OF CONTENTS4
FIG. 1: PULSE GENERATOR TEST ARRANGEMENT5
GENERAL OPERATING INSTRUCTIONS6
FIG. 2: DIODE SOCKET PIN CONNECTIONS7

Manual Reference: /fileserver1/officefiles/instructword/av-149/AV-149-BW4-2.8K-PIN-eda.doc, created July 14, 2000

FIG. 1: PULSE GENERATOR TEST ARRANGEMENT



- 1) The basic operation of the amplifier was confirmed using the pulse mode test arrangement shown in Fig. 1.
- 2) The AV-149 amplifier requires a prime power of +15 VDC (100 mA max).
- 3) <u>CAUTION</u>: The amplifier will be damaged if the anode and cathode connections are reversed.
- 4) The PIN diode bias may be varied from +1 to +14 Volts using the ten turn "BIAS" trim pot. The unit is shipped with the bias set at +14 Volts. The bias voltage may be monitored at the "BIAS" female banana connector.
- 5) The leads of the photo diode should not be inserted more than 1.0 cm into the socket.
- 6) The +15 VDC supply should be turned off when installing or removing the photo diode.
- 7) To test the AV-149 in a sweep frequency mode (10 MHz to 1200 MHz), the AVK-AV-C and the sampling scope should be replaced by a network analyser.
- 8) For additional information:

Tel:	(613) 226-5772
Fax:	(613) 226-2802
Email:	info@avtechpulse.com

FIG. 2: DIODE SOCKET PIN CONNECTIONS

- Anode
 Case (grounded)
 Cathode