

AVTECH ELECTROSYSTEMS LTD.

NANOSECOND WAVEFORM ELECTRONICS SINCE 1975

P.O. BOX 265 OGDENSBURG, NY U.S.A. 13669-0265 TEL: (315) 472-5270 FAX: (613) 226-2802 TEL: 1-800-265-6681 FAX: 1-800-561-1970

e-mail: info@avtechpulse.com http://www.avtechpulse.com P.O. BOX 5120 STN. F OTTAWA, ONTARIO CANADA K2C 3H4 TEL: (613) 226-5772 FAX: (613) 226-2802

INSTRUCTIONS

MODEL AV-141C1 AMPLIFIER

S.N.:

WARRANTY

Avtech Electrosystems Ltd. warrants products of its manufacture to be free from defects in material and workmanship under conditions of normal use. If, within one year after delivery to the original owner, and after prepaid return by the original owner, this Avtech product is found to be defective, Avtech shall at its option repair or replace said defective item. This warranty does not apply to units which have been dissembled, modified or subjected to conditions exceeding the applicable specifications or ratings. This warranty is the extent of the obligation assumed by Avtech with respect to this product and no other warranty or guarantee is either expressed or implied.

TECHNICAL SUPPORT

Phone: 613-226-5772 or 1-800-265-6681 Fax: 613-226-2802 or 1-800-561-1970

E-mail: info@avtechpulse.com
World Wide Web: http://www.avtechpulse.com

TABLE OF CONTENTS

WARRANTY	2
TABLE OF CONTENTS	3
MODEL AV-141C1	4
GENERAL INSTRUCTIONS	4
FIG. 1: AMPLIFIER TEST ARRANGEMENT	5

MODEL AV-141C1

GENERAL INSTRUCTIONS

The Model AV-141C1 amplifier is designed to amplify bipolar nanosecond rise time baseband pulses in the pulse width range of about 2.0 ns and higher and CW signals in the frequency range of DC to 800 MHz. The basic specifications for the unit are as follows:

Gain: > 20 dB

Peak output

voltage: ± 3 Volt

Rise time: < 700 ps

Impedance level: 50 Ohms nominal

Bandwidth: DC to 800 MHz

Max. noise figure: 4 nV/ Hz

Prime power: +15 Volts, 50 mA max


-15 Volts, 50 mA max


Connectors: SMA


Size: 1.4 x 1.1 x 2.3 inches

- 1) The DC offset on the output may be set to zero by making minor adjustments to the OS trim pot on the "OUT" end of the chassis.
- 2) The pulse mode operation of the unit was tested using the arrangement shown in Fig. 1.
- 2) CAUTION: The unit may be damaged if the DC voltage power supply voltage exceeds ±18.0 Volts or if the power supply voltages are reversed. The 1N4746A zener diodes are installed across the ±15 VDC input terminals to protect the unit against these factors. Note that the input power supply voltages may be as low as ±7 Volts.
- 4) <u>CAUTION</u>: The unit may be damaged if the input voltage exceeds ±300 mV, the input is not protected. Failures due to overvoltage application are not covered by the warranty.

FIG. 1: AMPLIFIER TEST ARRANGEMENT







6

model 2417.