

P.O. BOX 265
OGDENSBURG, NY
U.S.A. 13669-0265
TEL: (315) 472-5270
FAX: (613) 226-2802

AVTECH ELECTROSYSTEMS LTD.

NANOSECOND WAVEFORM ELECTRONICS SINCE 1975

TEL: 1-800-265-6681 FAX: 1-800-561-1970 U.S.A. & CANADA BOX 5120 STN. F OTTAWA, ONTARIO CANADA K2C 3H4 TEL: (613) 226-5772 FAX: (613) 226-2802

INSTRUCTIONS

MODEL AVR-A-1-PW-C-PN-EAD PULSE GENERATOR

S.N.:

WARRANTY

Electrosystems Avtech Ltd. warrants products of its in material manufacture to be free from defects and workmanship under conditions of normal use. If, within one year after delivery to the original owner, and after prepaid return by the original owner, this Avtech product is found to be defective, Avtech shall at its option repair or replace said defective item. This warranty does not apply to units been dissembled, modified or subjected to which have conditions exceeding the applicable specifications or ratings. This warranty is the extent of the obligation or liability assumed by Avtech with respect to this product and no other warranty or guarantee is either expressed or implied.

PULSE GENERATOR TEST ARRANGEMENT



Fig. l

Notes:

- The bandwidth capability of components and instruments used to display the pulse generator output signal (attenuators, cables, connectors, etc.) should exceed 100 MHz.
- 2) The use of 60 db attenuator at the scope vertical input channel will insure a peak input signal to the scope of less than one volt (necessary only if sampling scope used). If a high impedance real time scope is used, the pulse generator should be terminated using a shunt 50 ohm resistor.
- 3) The TRIG output channel provides TTL level signals. To avoid overdriving the TRIG input channel of some scopes, a 30 db attenuator should be placed at the input to the scope trigger channel. The TRIG output precedes the main output when the front panel ADVANCE-DELAY switch is in the ADVANCE position. The TRIG output lags the main output when the switch is in the DELAY position.
- 4) To obtain a stable output display the PW and PRF controls on the front panel should be set mid range. The front panel TRIG toggle switch should be in the INT position. The DELAY controls and the scope triggering controls are then adjusted to obtain a stable output. The scope may then be used to set the desired PRF by rotating the PRF controls.
- 4a) The output polarity is controlled by the front panel two position polarity switch.
- 5) The output pulse width is controlled by means of the front panel one turn PW control.
- 6) The output amplitude is controlled by means of the front panel one turn AMP control when the rear panel INT-EA-EAD switch is in the INT position. The control should initially be set maximum counter-clockwise and the amplitude adjusted using an oscilloscope. To voltage control the output amplitude, set the rear panel switch in the EA position and apply 0 to +10V between terminal A and ground $(R_{TN} > 10K)$. (EA option). To digitally control the output pulse width (in 256 increments) set the rear panel INT-EA-EAD switch in the EAD position and apply a parallel TTL control to PINS 1 to 8 (PIN 1 = LSB, PIN 8 = MSB) to the rear panel EWD D connector. PIN 14 is ground (EAD option).

- 7) An external clock may be used to control the output PRF of the AVR unit by setting the front panel TRIG toggle switch in the EXT position and applying a 0.2 us (approx.) TTL level pulse to the TRIG BNC connector input. For operation in this mode, the scope time base must also be triggered by the external clock.
- 8) The AVR-1-PW features an output impedance of the order of several ohms (rather than 50 ohms). The following consequences of this feature should be noted:
 - a) When used to switch some semiconductor devices (eg. bipolar and VMOS power transistors), the AVR unit will yield much faster switching times than those provided by 50 ohm pulse generators.
 - b) The AVR unit will safely operate in to load impedances in the range of 50 ohms to an open circuit. However, the fall time may degrade for load impedances higher than fifty ohms.
 - c) The AVR unit may be effectively converted to a fifty ohm output impedance generator by placing a fifty ohm 1/2 watt carbon composition resistor in series with the output of the unit and the load. The maximum available load voltage will then decrease to 100 volts (from 200 volts).
 - d) The output switching elements may fail if the unit is inadvertently operated into a short circuit. The switching elements are easily replaced in the field following the procedure outlined in the REPAIR Section.
- 9) The maximum allowable output pulse width for each PRF range (1, 2 and 3, MAX and MIN) is given in the following table. The output amplitude will decrease and the rear panel 1.0A slow blow fuse may blow and in extreme cases, the unit may fail if the pulse width (ie. duty cycle) conditions are exceeded.

		(usec)
Range 1		
PRF MAX	$(\approx 1.5 \text{ kHz})$	1 usec
PRF MIN	$(\approx 100 \text{ Hz})$	1 usec
Range 2		
PRF MAX	$(\approx 20 \text{ kHz})$	0.2 usec
PRF MIN	(≈1.5 kHz)	1 usec
Range 3		
PRF MAX	$(\approx 100 \text{ kHz})$	0.05 usec
PRF MIN	$(\approx 10 \text{ kHz})$	0.5 usec

MAX PW

- 10) AVR-A-1-PW-C units with a serial number higher than 5600 are protected by an automatic overload protective circuit which controls the front panel overload light. If the unit is overloaded (by operating at an exceedingly high duty cycle or by operating into a short circuit), the protective circuit will turn the output of the instrument OFF and turn the indicator light ON. The light will stay ON (i.e. output OFF) for about 5 seconds after which the instrument will attempt to turn ON (i.e. light OFF) for about 1 second. If the overload condition persists, the instrument will turn OFF again (i.e. light ON) for another 5 seconds. If the overload condition has been removed, the instrument will turn on and resume normal operation. Overload conditions may be removed by:
 - 1) Reducing PRF (i.e. switch to a lower range)
 - 2) Reducing pulse width (i.e. switch to a lower range)
 - 3) Removing output load short circuit (if any)
- 11) The unit can be converted from 110 to 220V 50-60 Hz operation by adjusting the voltage selector card in the rear panel fused voltage selector-cable connector assembly.
- 12) For additional assistance:

Tel: (613) 226-5772 Fax: (613) 226-2802

- the sum and the state of the ------4 % -<u>|</u>--· 1 ر بالقدرية. 1. ÷ · ···· 72. · · · · -1.0 5 322 :3 -0.9 MAXIMUM_PW VERSUS PRF FOR MODEL MR-A-1 (R_= 50_ Eig_ 0.8 -0.1 PW RAX 04 (11SEZ) -----20 0.4 0.3 0.2-0.-1-0 5 10 40. 50 30 -1 60 80 90 20 100

.

Fig. 2



•

Ì.

- (1) <u>ON-OFF Switch</u>. Applies basic prime power to all stages.
- (2) <u>PRF Control</u>. Controls PRF as follows:

Range	1	10	Hz	to	100	Hz
Range	2	100	Hz	to	1	kHz
Range	3	1	kHz	to	10	kHz
Range	4	10	kHz	to	100	kHz

- (3) <u>DELAY Control</u>. Controls the relative delay between the reference output pulse provided at the TRIG output (4) and the main output (5). This delay is variable over the range of 0 to about 1.0 us. The TRIG output precedes the main output when the ADVANCE-DELAY switch is in the ADVANCE position and lags when the switch is in the DELAY position.
- (4) <u>TRIG Output</u>. This output is used to trigger the scope time base. The output is a TTL level 100 ns (approx.) pulse capable of driving a fifty ohm load.
- (5) <u>OUT Connector</u>. BNC connector provides output to a fifty ohm load.
- (6) <u>PW Control</u>. A one turn control which varies the output pulse width from 0.05 us to 1.0 us.
- (7) <u>AMP Control</u>. A one turn control which varies the output pulse amplitude from 0 to 200V to a fifty ohm load.
- (8) <u>EXT-INT Control</u>. With this toggle switch in the INT position, the PRF of the AVR unit is controlled via an internal clock which in turn is controlled by the PRF controls. With the toggle switch in the EXT position, the AVR unit requires a 50 ns TTL level pulse applied at the TRIG input in order to trigger the output stages. In addition, in this mode, the scope time base must be triggered by the external trigger source.

BACK PANEL CONTROLS



- (1) <u>FUSED CONNECTOR, VOLTAGE SELECTOR</u>. The detachable power cord is connected at this point. In addition, the removable cord is adjusted to select the desired input operating voltage. The unit also contains the main power fuse (0.50 A SB).
- (2) <u>1.0 A SB</u>. Fuse which protects the output stage if the output duty cycle rating is exceeded.
- (3) To voltage control the output amplitude, set the switch
- (4) (3) in the EA position and apply 0 to +10V between the
- (5) terminal A (4) and ground ($R_{IN} \ge 10K$). (EA option).

EAD option: To digitally control the output amplitude (in 256 increments) set the switch (3) in the EAD position and apply a parallel TTL control to PINS 1 to 8 (PIN 1 = LSB, PIN 8 = MSB) to the EAD D connector (5). PIN 24 is ground.

Logic Level		Volts	Current
0	0	to +0.8V	10 uA (max)
1	+2	to +5V	10 uA (max)





SYSTEM DESCRIPTION AND REPAIR PROCEDURE

The AVR-A-1-C consists of the following basic modules:

- 1) AVR-A-1-PG pulse generator module
- 2) AVR-A-1-CL clock module
- 3) +24V power supply board

The modules are interconnected as shown in Fig. 4.

The clock module controls the output PRF and the relative delay between the main output and the SYNC outputs. The PG pulse generator module generates the output pulse. In the event of an instrument malfunction, it is most likely that some of the output switching elements (SL4) may have failed due to an output short circuit condition or to a high duty cycle condition. The switching elements may be accessed by removing the cover plate on the bottom side of the First turn off the prime power. The instrument. NOTE: elements may be removed from their sockets by means of a needle nosed pliers. The SL4 is a selected VMOS power transistor in a TO 220 packages and may be checked on a curve If defective, replacement units should be ordered tracer. directly from Avtech. When replacing the SL4 switching elements, take care to insure that the short lead (of the three leads) is adjacent to the black dot on the chassis. If the switching elements are not defective, then the four Phillips screws on the back panel should be removed. The top cover may then be slid off and operation of the clock and power supply modules should be checked. The clock module is functioning properly if:

- a) 0.1 usec TTL level outputs are observed at pins 2 and 3.
- b) The PRF of the outputs can be varied over the range of 0.1 KHz to 0.1 MHz using the PRF and PRF FINE controls.
- c) The relative delay between the pin 2 and 3 outputs can be varied by at least 500 nsec by the DELAY controls.

The sealed clock module must be returned to Avtech for repair or replacement if the above conditions are not observed. The power supply board generates +24V DC to power the other modules. If the voltage is less than +24V, turn off the prime power and unsolder the lead from the 7824 regulator chip on the power supply board. Solder a 100 ohm 5 watt resistor to the 7824 output to ground and turn on the prime power. A voltage of +24 volts should be read. If the voltage is less then the power supply board is defective and should be repaired or replaced.

March 9/94

- PN

- FAD BED DATE TRANSPORTER AND FOR AND FOR A

teratobole relatives for any relatives. I transverself

The are faired and a connect of an shown on Fig. 3.

We close notice control to be output to not to the notice 100° not on the control to the number of the second to the the second to the s

(4) You's set that there is define and a case of the energy state of the state o