



AVTECH ELECTROSYSTEMS LTD.

NANOSECOND WAVEFORM ELECTRONICS
SINCE 1975

info@avtechpulse.com
<http://www.avtechpulse.com/>

Tel: 888-670-8729 (USA & Canada)
or +1-613-686-6675 (Worldwide)

BOX 5120, LCD MERIVALE
OTTAWA, CANADA K2C3H5

INSTRUCTIONS

MODEL AV-112D-PS

0 to +50 V, 20 kHz

VARIABLE-GAIN

LINEAR AMPLIFIER

SERIAL NUMBER: 14135

WARRANTY

Avtech Electrosystems Ltd. warrants products of its manufacture to be free from defects in material and workmanship under conditions of normal use. If, within one year after delivery to the original owner, and after prepaid return by the original owner, this Avtech product is found to be defective, Avtech shall at its option repair or replace said defective item. This warranty does not apply to units which have been disassembled, modified or subjected to conditions exceeding the applicable specifications or ratings. This warranty is the extent of the obligation assumed by Avtech with respect to this product and no other warranty or guarantee is either expressed or implied.

TECHNICAL SUPPORT

Phone: 888-670-8729 (USA & Canada)
or +1-613-686-6675 (Worldwide)

E-mail: info@avtechpulse.com
World Wide Web: <http://www.avtechpulse.com>

TABLE OF CONTENTS

WARRANTY	2
TECHNICAL SUPPORT	2
TABLE OF CONTENTS	3
INTRODUCTION	5
SPECIFICATIONS	6
REGULATORY NOTES	7
FCC PART 18.....	7
EC DECLARATION OF CONFORMITY.....	7
DIRECTIVE 2011/65/EU (RoHS).....	8
DIRECTIVE 2002/96/EC (WEEE).....	8
INSTALLATION	10
VISUAL CHECK.....	10
POWER RATINGS.....	10
CONNECTION TO THE POWER SUPPLY.....	10
PROTECTION FROM ELECTRIC SHOCK.....	11
ENVIRONMENTAL CONDITIONS.....	12
FUSES	13
AC FUSE REPLACEMENT.....	13
DC FUSE REPLACEMENT.....	14
FUSE RATINGS.....	14
FRONT PANEL CONTROLS	15
REAR PANEL CONTROLS	16
GENERAL INFORMATION	17
BASIC CONTROL.....	17
POWER ON / OFF.....	17
PROTECTING THE INSTRUMENT.....	17
OPERATIONAL CHECK	19
MECHANICAL INFORMATION	21
TOP COVER REMOVAL.....	21
RACK MOUNTING.....	21
ELECTROMAGNETIC INTERFERENCE.....	21
MAINTENANCE	22
REGULAR MAINTENANCE.....	22
CLEANING.....	22

WIRING DIAGRAMS.....23

- WIRING OF AC POWER.....23**
- HIGH VOLTAGE POWER SUPPLY.....24**
- PCB 158R5 - LOW VOLTAGE POWER SUPPLY.....25**
- PCB 161C - DC/DC CONVERTERS.....26**
- PCB 94H - ALARM BOARD.....27**
- PCB 288B - POWER DISSIPATION LIMITER.....28**
- PCB 240A - OUTPUT RELAY.....29**
- PCB 183A-S AND 183A-P CAPACITOR BANKS.....30**
- MAIN WIRING.....31**

PERFORMANCE CHECK SHEET.....32

Manual Reference: /fileserver2/officefiles/instructword/av-112/AV-112D-PS,ed2.odt.
Last modified June 15, 2021.
Copyright © 2021 Avtech Electrosystems Ltd, All Rights Reserved.

INTRODUCTION

The Model AV-112D-PS variable-gain linear amplifier accepts input voltages in the range of 0 to +2V, and has a variable gain of 0 to +25. The maximum output voltage is +50V.

The AV-112D-PS will drive load impedances of 10 Ω or higher. The maximum peak output current is 5 Amps (that is, 50 V / 10 Ω = 5 A). The maximum average current is 2.5 Amps (measured over a 10 ms interval). The maximum average power that can be delivered to a load is 125 W. The instrument is protected against overload conditions. However, the user should ensure that the load impedance is never less than 10 Ω , to avoid over-stressing the output circuitry. The input is protected against excessive voltages.

This instrument is intended for use in research, development, test and calibration laboratories by qualified personnel.

SPECIFICATIONS

Model:	AV-112D-PS
Output amplitude:	0 to +50 Volts
Maximum peak current:	5 Amps
Maximum average current ⁵ :	2.5 Amps
Output polarity:	Positive only
Load impedance:	$\geq 10 \Omega$
Output resistance ¹ :	$\approx 0 \Omega$
Bandwidth (f_{-3dB}):	20 kHz
Rise time (20%-80%, for maximum output):	3 us
Average output power: (maximum)	125 W
Voltage gain ² :	$\times 1$ to $\times 25$
Gain polarity:	Non-inverting (+)
Input range ² :	0 to +2 Volts
Input impedance:	1 k Ω
DC offset option ³ :	0 to +35 Volts
Connectors:	In, Out: BNC
Dimensions:	100 mm x 430 mm x 375 mm (3.9" x 17" x 14.8")
Power requirement:	100-240 Volts, 50-60 Hz

- 1) "Output resistance" is the internal resistance in series with output. Non-zero output impedances (R_{OUT}) will reduce the maximum output amplitude slightly when operating into low load impedances. That is, $V_{OUT} = V_{SET} \times R_{LOAD} / (R_{LOAD} + R_{OUT})$, where V_{SET} is the set amplitude and R_{LOAD} is the load resistance.
- 2) These parameters can easily be adapted to meet special requirements. Contact Avtech (info@avtechpulse.com) with your special application.
- 3) To specify the DC offset option, add the suffix -OS to the model number. The sum of the amplitude and the offset must remain within the rated output amplitude range – i.e., this option does not change the minimum or maximum obtainable output voltage.
- 4) Add the suffix -INV to specify the switchable gain polarity feature.
- 5) Averaged over a 10 ms interval.

REGULATORY NOTES

FCC PART 18

This device complies with part 18 of the FCC rules for non-consumer industrial, scientific and medical (ISM) equipment.

This instrument is enclosed in a rugged metal chassis and uses a filtered power entry module (where applicable). The main output signal is provided on a shielded connector that is intended to be used with shielded coaxial cabling and a shielded load. Under these conditions, the interference potential of this instrument is low.

If interference is observed, check that appropriate well-shielded cabling is used on the output connectors. Contact Avtech (info@avtechpulse.com) for advice if you are unsure of the most appropriate cabling. Also, check that your load is adequately shielded. It may be necessary to enclose the load in a metal enclosure.

If any of the connectors on the instrument are unused, they should be covered with shielded metal "dust caps" to reduce the interference potential.

This instrument does not normally require regular maintenance to minimize interference potential. However, if loose hardware or connectors are noted, they should be tightened. Contact Avtech (info@avtechpulse.com) if you require assistance.

EC DECLARATION OF CONFORMITY



We Avtech Electrosystems Ltd.
 P.O. Box 5120, LCD Merivale
 Ottawa, Ontario
 Canada K2C 3H5

declare that this pulse generator meets the intent of Directive 2004/108/EG for Electromagnetic Compatibility. Compliance pertains to the following specifications as listed in the official Journal of the European Communities:

EN 50081-1 Emission

EN 50082-1 Immunity

and that this pulse generator meets the intent of the Low Voltage Directive 2006/95/EC. Compliance pertains to the following specifications as listed in the official Journal of the European Communities:

EN 61010-1:2001 Safety requirements for electrical equipment for measurement, control, and laboratory use

DIRECTIVE 2011/65/EU (RoHS)

We Avtech Electrosystems Ltd.
P.O. Box 5120, LCD Merivale
Ottawa, Ontario
Canada K2C 3H5

declare that, to the best of our knowledge, all electrical and electronic equipment (EEE) sold by the company are in compliance with Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (also known as “RoHS Recast”). In addition, this declaration of conformity is issued under the sole responsibility of Avtech Electrosystems Ltd. Specifically, products manufactured do not contain the substances listed in the table below in concentrations greater than the listed maximum value.


<i>Material/Substance</i>	<i>Threshold level</i>
Lead (Pb)	< 1000 ppm (0.1% by mass)
Mercury (Hg)	< 1000 ppm (0.1% by mass)
Hexavalent Chromium (Cr6+)	< 1000 ppm (0.1% by mass)
Polybrominated Biphenyls (PBB)	< 1000 ppm (0.1% by mass)
Polybrominated Diphenyl ethers (PBDE)	< 1000 ppm (0.1% by mass)
Cadmium (Cd)	< 100 ppm (0.01% by mass)
Bis(2-ethylhexyl) phthalate (DEHP)	< 1000 ppm (0.1% by mass)
Butyl benzyl phthalate (BBP)	< 1000 ppm (0.1% by mass)
Dibutyl phthalate (DBP)	< 1000 ppm (0.1% by mass)
Diisobutyl phthalate (DIBP)	< 1000 ppm (0.1% by mass)

DIRECTIVE 2002/96/EC (WEEE)

European customers who have purchased this equipment directly from Avtech will have completed a “WEEE Responsibility Agreement” form, accepting responsibility for WEEE compliance (as mandated in Directive 2002/96/EC of the European Union and local laws) on behalf of the customer, as provided for under Article 9 of Directive 2002/96/EC.

Customers who have purchased Avtech equipment through local representatives should consult with the representative to determine who has responsibility for WEEE compliance. Normally, such responsibilities will lie with the representative, unless other arrangements (under Article 9) have been made.

Requirements for WEEE compliance may include registration of products with local governments, reporting of recycling activities to local governments, and financing of recycling activities.



INSTALLATION

VISUAL CHECK

After unpacking the instrument, examine to ensure that it has not been damaged in shipment. Visually inspect all connectors, knobs, and handles. Confirm that a power cord is with the instrument. If the instrument has been damaged, file a claim immediately with the company that transported the instrument.

POWER RATINGS


This instrument is intended to operate from 100 - 240 V, 50 - 60 Hz.

The maximum power consumption is 220 Watts. Please see the “FUSES” section for information about the appropriate AC and DC fuses.

This instrument is an “Installation Category II” instrument, intended for operation from a normal single-phase supply.

CONNECTION TO THE POWER SUPPLY


An IEC-320 three-pronged recessed male socket is provided on the back panel for AC power connection to the instrument. One end of the detachable power cord that is supplied with the instrument plugs into this socket. The other end of the detachable power cord plugs into the local mains supply. Use only the cable supplied with the instrument. The mains supply must be earthed, and the cord used to connect the instrument to the mains supply must provide an earth connection. (The supplied cord does this.)

 Warning: Failure to use a grounded outlet may result in injury or death due to electric shock. This product uses a power cord with a ground connection. It must be connected to a properly grounded outlet. The instrument chassis is connected to the ground wire in the power cord.

The table below describes the power cord that is normally supplied with this instrument, depending on the destination region:

Destination Region	Description	Option	Manufacturer	Part Number
United Kingdom, Hong Kong, Singapore, Malaysia	BS 1363, 230V, 50 Hz	-AC00	Qualtek	370001-E01
Australia, New Zealand	AS 3112:2000, 230-240V, 50 Hz	-AC01	Qualtek	374003-A01
Continental Europe, Korea, Indonesia, Russia	European CEE 7/7 "Schuko" 230V, 50 Hz	-AC02	Qualtek	364002-D01
North America, Taiwan	NEMA 5-15, 120V, 60 Hz	-AC03	Qualtek	312007-01
Switzerland	SEV 1011, 230V, 50 Hz	-AC06	Qualtek	378001-E01
South Africa, India	SABS 164-1, 220-250V, 50 Hz	-AC17	Volex	2131H 10 C3
Japan	JIS 8303, 100V, 50-60 Hz	-AC18	Qualtek	397002-01
Israel	SI 32, 220V, 50 Hz	-AC19	Qualtek	398001-01
China	GB 1002-1 / 2099-1, 220V, 50 Hz	-AC22	Qualtek	399012-01

PROTECTION FROM ELECTRIC SHOCK

 Operators of this instrument must be protected from electric shock at all times. The owner must ensure that operators are prevented access and/or are insulated from every connection point. In some cases, connections must be exposed to potential human contact. Operators must be trained to protect themselves from the risk of electric shock. This instrument is intended for use by qualified personnel who recognize shock hazards and are familiar with safety precautions required to avoid possibly injury. In particular, operators should:

1. Keep exposed high-voltage wiring to an absolute minimum.
2. Wherever possible, use shielded connectors and cabling.
3. Connect and disconnect loads and cables only when the instrument is turned off.
4. Keep in mind that all cables, connectors, oscilloscope probes, and loads must have an appropriate voltage rating.
5. Do not attempt any repairs on the instrument, beyond the fuse replacement procedures described in this manual. Contact Avtech technical support (see page 2 for contact information) if the instrument requires servicing. Service is to be performed solely by qualified service personnel.


ENVIRONMENTAL CONDITIONS

This instrument is intended for use under the following conditions:

1. indoor use;
2. altitude up to 2 000 m;
3. temperature 5 °C to 40 °C;
4. maximum relative humidity 80 % for temperatures up to 31 °C decreasing linearly to 50 % relative humidity at 40 °C;
5. Mains supply voltage fluctuations up to ± 10 % of the nominal voltage;
6. no pollution or only dry, non-conductive pollution.


FUSES

This instrument contains four fuses. All are accessible from the rear-panel. Two protect the AC prime power input, and two protect the internal DC power supplies. The locations of the fuses on the rear panel are shown in the figure below:



AC FUSE REPLACEMENT

To physically access the AC fuses, the power cord must be detached from the rear panel of the instrument. The fuse drawer may then be extracted using a small flat-head screwdriver, as shown below:



DC FUSE REPLACEMENT

The DC fuses may be replaced by inserting the tip of a flat-head screwdriver into the fuse holder slot, and rotating the slot counter-clockwise. The fuse and its carrier will then pop out.

FUSE RATINGS


The following table lists the required fuses:

Fuses	Nominal Mains Voltage	Rating	Case Size	Recommended Replacement Part	
				Littelfuse Part Number	Digi-Key Stock Number
#1, #2 (AC)	115 V	2.0A, 250V, Time-Delay	5×20 mm	0218002.HXP	F2420-ND
	230 V	1.0A, 250V, Time-Delay	5×20 mm	0218001.HXP	F2419-ND
#3 (DC)	N/A	0.5A, 250V, Time-Delay	5×20 mm	0218.500HXP	F2416-ND
#4 (DC)	N/A	10A, 250V, Time-Delay	5×20 mm	0218010.HXP	F2425-ND

The recommended fuse manufacturer is Littelfuse (<http://www.littelfuse.com>).

Replacement fuses may be easily obtained from Digi-Key (<http://www.digikey.com>) and other distributors.


FRONT PANEL CONTROLS




1. POWER Switch. This is the main power switch. When turning the instrument on, there may be a delay of several seconds before the instrument appears to respond.
2. OVERLOAD Indicator. When the instrument is powered, this indicator is normally green, indicating normal operation. If this indicator is yellow, an internal automatic overload protection circuit has been tripped. If the unit is overloaded (by operating at an exceedingly high duty cycle or by operating into a very low impedance), the protective circuit will disable the output of the instrument and turn the indicator light yellow. The light will stay yellow (i.e. output disabled) for about 5 seconds after which the instrument will attempt to re-enable the output (i.e. light green) for about 1 second. If the overload condition persists, the output will be disabled again (i.e. light yellow) for another 5 seconds. If the overload condition has been removed, the instrument will resume normal operation.

This overload indicator is only likely to come on in two situations:

- Briefly at startup. This is not a cause for concern.
 - When the load impedance is too low ($< 10 \Omega$). In this case, turn off the instrument and connect the proper load.
3. IN Connector. The input signal is applied to this connector. The input impedance is approximately 1 k Ω . The input should range between 0 and +2 Volts..
 4. GAIN Dial. This ten-turn dial is used to vary the amplifier gain between 0 and +25.
 5. OUT Connector. This BNC connector provides the main output signal. The output is an amplified version of the input on (3). The gain (V_{OUT}/V_{IN}) is controlled by (4).

 **Caution:** Voltages as high as +50V may be present on the center conductor of this output connector. Avoid touching this conductor. Connect to this connector using standard coaxial cable, to ensure that the center conductor is not exposed.

REAR PANEL CONTROLS



1. AC POWER INPUT. An IEC-320 C14 three-pronged recessed male socket is provided on the back panel for AC power connection to the instrument. One end of the detachable power cord that is supplied with the instrument plugs into this socket.
2. AC FUSE DRAWER. The two fuses that protect the AC input are located in this drawer. Please see the “FUSES” section of this manual for more information.
3. DC FUSES. These two fuses protect the internal DC power supplies. Please see the “FUSES” sections of this manual for more information.


GENERAL INFORMATION

BASIC CONTROL

The AV-112D-PS is a DC-20 kHz variable-gain linear amplifier. The gain is variable from 0 to +25, and is adjusted by rotating the "GAIN" control.

The required voltage input signal is applied at the "IN" connector.

This mode is illustrated below:



POWER ON / OFF

The instrument has a 5 second warm-up time after switching the power on. During this time, the amplifier may function, but the maximum output will be limited in amplitude as the internal high-voltage power supply charges up.

Unpredictable transients may occur on the output during a two-second interval after switching the instrument off. If your load is sensitive to such transients, disconnect your load before switching off the power.

The instrument may beep several times, and the overload indicator may flash, during power-off. This is normal, and not a cause for concern.

PROTECTING THE INSTRUMENT


Always ensure that the equivalent load resistance (V_{OUT} / I_{OUT}) is 10 Ohms or greater.

The output is thermally protected, and the average output power is monitored and limited. However, sudden short circuit conditions in the load may damage the output circuitry. Take care to avoid short circuit conditions!

The input is protected against excessive voltages.

OPERATIONAL CHECK

This section describes a sequence to confirm the basic operation of the instrument. It should be performed after receiving the instrument. It is a useful learning exercise as well.



Basic Test Arrangement


- 1) Connect a 10 Ω , 125 W non-inductive test load between the OUT connector and ground. A higher resistance may also be used, in which case the power rating may be reduced. If cabling is used, keep it less than 3 feet / 1 meter in length. Confirm that the oscilloscope and the test load are rated for 50 Volt operation.
- 2) Set the signal generator to produce a +2V, 1 kHz waveform with a duty cycle of less than 50%. (The input impedance of the AV-112D-PS is 1 k Ω). Connect a cable from the SYNC connector of the signal generator to the TRIG input of an oscilloscope. Set the oscilloscope to trigger externally. Connect the main output of the signal generator to the input of the amplifier.
- 3) Connect one oscilloscope probe (channel A) to the output of the signal generator. Set the Channel A vertical scale to 1 V/div.
- 4) Connect one oscilloscope probe (channel B) to the 10 Ω load. On the oscilloscope, set the channel A vertical scale to 20 V/div, and the horizontal scale to 1 ms/div.

- 5) Set the gain control to minimum (0.0). Turn on the amplifier and the signal generator.
- 6) Channel A should show a +2V, 1 kHz waveform from the signal generator. Rotate the gain control to its maximum setting. The Channel B waveform should increase to +50V, and have a shape similar to that of the Channel A waveform.
- 7) This completes the operational check.


MECHANICAL INFORMATION

TOP COVER REMOVAL

If necessary, the interior of the instrument may be accessed by removing the four Phillips screws on the top panel. With the four screws removed, the top cover may be slid back (and off).

 Always disconnect the power cord and allow the instrument to sit unpowered for 10 minutes before opening the instrument. This will allow any internal stored charge to discharge.

There are no user-adjustable internal circuits. For repairs other than fuse replacement, please contact Avtech (info@avtechpulse.com) to arrange for the instrument to be returned to the factory for repair. Service is to be performed solely by qualified service personnel.

 Caution: High voltages are present inside the instrument during normal operation. Do not operate the instrument with the cover removed.

RACK MOUNTING

A rack mounting kit is available. The -R5 rack mount kit may be installed after first removing the one Phillips screw on the side panel adjacent to the front handle.

ELECTROMAGNETIC INTERFERENCE

To prevent electromagnetic interference with other equipment, all used outputs should be connected to shielded loads using shielded coaxial cables. Unused outputs should be terminated with shielded coaxial terminators or with shielded coaxial dust caps, to prevent unintentional electromagnetic radiation. All cords and cables should be less than 3m in length.

MAINTENANCE

REGULAR MAINTENANCE

This instrument does not require any regular maintenance.


On occasion, one or more of the four rear-panel fuses may require replacement. All fuses can be accessed from the rear panel. See the “FUSES” section for details.

CLEANING


If desired, the interior of the instrument may be cleaned using compressed air to dislodge any accumulated dust. (See the “TOP COVER REMOVAL” section for instructions on accessing the interior.) No other cleaning is recommended.

WIRING OF AC POWER

WIRING DIAGRAMS




HIGH VOLTAGE POWER SUPPLY




AV-112D-FS HIGH VOLTAGE FS	
Printed	15 Jun 2021
Revision	2C
Z:\engineering\circuits\av-11x-15\FUNC.ddb - AV-112D-FS-112D-HV-V2.sch	

PCB 158R5 - LOW VOLTAGE POWER SUPPLY



- USAGE NOTES:**
- PN 125W UVs NEED AT LEAST 0.22 OHMS OF SERIES RESISTANCE (EACH) TO TAME START-UP CURRENTS FOR BITS TO START UP HAPPILY.
 - MIN PRACTICAL LIMIT CURRENT:** 200K TO TRIP AROUND 0.35A. CAN NOT SET TRIP CURRENT BELOW 0.3A.
- 158R2 MODS:**
- DONE: BETTER POWER SEQUENCING OF U6, U7
 - DONE: RIN AND RSHUNT ON B55138
 - DONE: MAKE U6, U7 POLYGRONS DIRECT CONNECT
 - DONE: L1-L4, L6, L8 PAD SPACING WRONG
 - DONE: SOP4 FOOTPRINT TOO TIGHT
- 158R3 MODS:**
- FIX FB52 FOOTPRINT
- 158R4 MODS:**
- ADD U13 (CUI PQA50, XP JTL60) + L/C/J
 - USE 785R-5/1.5-W36-C AS 7805 ALTERNATIVE
- 158R4b MOD:**
- WIDEN 78SR HOLES
- 158R5 MOD:**
- MORE SMT-FRIENDLY
 - REMOVED LED

PCB 161C - DC/DC CONVERTERS




USAGE CHART

FOR 101X, FLOATING +24V	FOR 155C, TO GENERATE -15V	WISC, TO GENERATE -24V
USE FDC68-24D 12	USE FDC68-24S 15	USE FDC68-24D 12
FILT+ COM UNFILT- UNFILT+ FILT-	TO FLOAT+ NIC TO FLOAT- NIC NIC	NIC SHORT TO UNFILT- SHORT TO COM TO GND -15V OUT NIC NIC TO GND -24V OUT

Title	
Date	Revision
4-NOV-2008	
Z:\m\jcf\files\pcb\161\FDC MOUNTING.ddb - Documents\161C\PCB161C.SCH	

PCB 94H - ALARM BOARD




CORR1:
 - ADD U2 VCC/GND
 - BUZZER IS POLARIZED
 - UPDATE PCB!

TODO:
 - NON-SMA OUTPUT FOR 106/107
 - INTERLOCK SMA AND PULLUP RES?

Sheet: /
 File: pcb94H_corr1.sch

Title:
 Size: A3 Date: Rev:
 KiCad E.D.A. eeschema 5.1.6-1.fc32 Id: 1/1

PCB 288B - POWER DISSIPATION LIMITER



ORIGINAL 108F-3 CALCCS:

IF R1,R3 = 20K
 $V734 = VMON * ((HV - VOUT)/3) / 10V$
 $-VMON = IOUT * 0.02 OHMS$
 $V734 = IOUT * 0.02 * (HV - VOUT) / 30$
 $V734 = PDISS / 1500$
 $80W / 1500 = 53 mV$
 - APPROX - IGNORES 50K DIFF INPUT RES OF AD734.
 USE R7
 = 59K FOR 50mV TRIP
 = 13.7K FOR 150mV TRIP

BETTER CALCS. AV-108F-4-B (S/N 13904):

R1, R3 = 100K (FOR 100V MAX, GIVES 11:1 DIV IN THEORY)
 BUT 50K ZIN REDUCES GAIN, GIVING 6V ACROSS Y1-Y2
 REMOVE R11/12, AND INSTALL R14 = 0 FOR DENOMINATOR CONTROL (MORE GAIN)
 INSTALL R13 = 300K, GIVES U1-U2 OF 1.2V APPROX (INSTEAD OF FIXED 10V)
 $WITH RMON = 0.005 OHM X 4 (GAIN ON 116) = 0.02 OHMS$
 10A gives 200 mV
 100V gives 6V
 $200 mV * 6V / 1.2V = 1V$
 SO 1V PER 1000W
 150mV = 150W
 USE R7 = 13.7K FOR 150mV TRIP AT 150W AVG

AV-109F-4-B (S/N 13938):


R1, R3 = 30K (FOR 10V MAX, GIVES 4:1 DIV IN THEORY)
 $WITH RMON = 0.22/30 X 5.6 (GAIN ON 116) = 0.04 OHMS$
 100A gives 4V
 10V gives 2.5V
 $4V * 2.5V / 10V = 1V$
 SO 1V PER 1000W
 150mV = 150W
 USE R7 = 13.7K FOR 150mV TRIP AT 150W AVG
 AS STARTING POINT

AV-112D-PS:

R1, R3 = 51K (FOR 60V MAX, GIVES 6:1 DIV IN THEORY,
 BUT RIN REDUCES GAIN BY HALF)
 WANT 60V DIFF GIVES 10V OUT
 $IMON = 1V/A (SO 2.5A GIVES 2.5V)$
 WANT TRIP AT 150W, SO TRIP AT V734 = +2.5V
 USE R7 = 470.2W (LEADED) AS STARTING POINT?
 IN PRACTICE, R7 = 1K WORKS BETTER (DUE TO RIN).

Sheet: /	
File: pcb288B.kicad_sch	
Title: PDISS PROTECTION	
Size: A3	Date: 2021-06-15
KiCad E.D.A. eeschema 5.99.0-1.20210608gited0e0a0.fc34	Id: 1/1

PCB 240A - OUTPUT RELAY



ORIGINAL 108F-3 CALCS:

IF R1,R3 = 20K

$$V734 = VMON * ((HV - VOUT)/3) / 10V$$

$$-VMON = IOUT * 0.02 OHMS$$

$$V734 = IOUT * 0.02 * (HV - VOUT) / 30$$

$$V734 = PDISS / 1500$$

$$80W / 1500 = 53 mV$$

- APPROX - IGNORES 50K DIFF INPUT RES OF AD734.

USE R7

= 59K FOR 50mV TRIP.

= 13.7K FOR 150mV TRIP

BETTER CALCS. AV-108F-4-B (S/N 13904):

R1, R3 = 100K (FOR 100V MAX, GIVES 11:1 DIV IN THEORY)
BUT 50K ZIN REDUCES GAIN, GIVING 6V ACROSS Y1-Y2

REMOVE R11/12, AND INSTALL R14 = 0 FOR DENOMINATOR CONTROL (MORE GAIN)
INSTALL R13 = 300K, GIVES U1-U2 OF 1.2V APPROX (INSTEAD OF FIXED 10V)

WITH RMON = 0.005 OHM X 4 (GAIN ON 116) = 0.02 OHMS

10A gives 200 mV

100V gives 6V

200 mV * 6V / 1.2V = 1V

SO 1V PER 1000W

150mV = 150W

USE R7 = 13.7K FOR 150mV TRIP AT 150W AVG

AV-109F-4-B (S/N 13938):

R1, R3 = 30K (FOR 10V MAX, GIVES 4:1 DIV IN THEORY)

WITH RMON = 0.22/30 X 5.6 (GAIN ON 116) = 0.04 OHMS

100A gives 4V

10V gives 2.5V

4V * 2.5V / 10V = 1V

SO 1V PER 1000W

150mV = 150W

USE R7 = 13.7K FOR 150mV TRIP AT 150W AVG

AS STARTING POINT

AV-112D-PS:

R1, R3 = 51K (FOR 60V MAX, GIVES 6:1 DIV IN THEORY,
BUT RIN REDUCES GAIN BY HALF)

WANT 60V DIFF GIVES 10V OUT

IMON = 1V/A (SO 2.5A GIVES 2.5V)

WANT TRIP AT 150W, SO TRIP AT V734 = +2.5V

USE R7 = 470.2W (LEADED) AS STARTING POINT?

IN PRACTICE, R7 = 1K WORKS BETTER (DUE TO RIN).

Sheet: /

File: pcb288B.kicad_sch

Title: PDISS PROTECTION

Size: A3

Date: 2021-06-15


Rev:

KiCad E.D.A. eeschema 5.99.0-1.20210608gited0e0a0.fc34




Id: 1/1

PCB 183A-S AND 183A-P CAPACITOR BANKS

183A-S (SERIES CAPACITOR BANK)



183A-P (PARALLEL CAPACITOR BANK)




PCB183 CAP BANKS

Printed 19-Aug-2005

Revision 1C

Z:\mjrfiles\pcb183\hv-cap-bank.Dtb - pcb183a.sch

MAIN WIRING



AV-112D-PS, HIGH-VOLTAGE FS		
Printed	15 Jun 2021	Revision 3B
Z:\engineering\circuit\av-11x-15\FUNC.dcb - AV-112D-PS-112Dwiring V3.sch		

PERFORMANCE CHECK SHEET